

RoyIntelli Technology Co., Ltd.
RI-C5800 | Automatic Bomb Calorimeter
Category: Calorific Value
Buyer datasheet | Version 1.0 | 2026-09-24

Purpose and sample types

- Coal, coke, oil, biomass, SRF
- ISO 1928:2009; ASTM D5865-13; GB/T 213-2008; CEN/TS 14918; BS EN 15400; JJG 672-2018
- Classic test ≤ 16 min; fast mode ≤ 11 min
- Temperature resolution 0.0001 C; precision RSD $\leq 0.15\%$; AC 220 V, 50 Hz
- Automatic oxygen filling, bomb lifting, water dosing and jacket control
- Factory documents, English manual and remote commissioning are available according to the confirmed order

Ordering checklist

- Confirm sample matrix, daily throughput, reporting standard, voltage and plug format.
- Request the model-specific quotation, acceptance protocol and available evidence pack.
- Final specifications, documents, lead time, warranty and payment terms are confirmed in the Proforma Invoice.

Documents and support

- English operating documentation and remote commissioning are available according to the confirmed order.
- CRM reports, calibration records, conformity documents and reference material are supplied or made available only when applicable and verifiable for the ordered model.

Contact

service@royintelli.com | +86 17808359107 | WhatsApp +852 57166955
www.royintelli.com | Changsha, Hunan, China

This datasheet is a procurement summary, not a certificate. Confirm method compatibility and all commercial details before purchase.